

MOUNTED

747621

JAN 11 1961

Temp. Coef. = — Microvolts per °C

B. S. Certified Value

1.01 Volts at °C

Eppley Certified Value \$117.63

1.01923 Abs. Volts at 24 °C

Shipped to Braun-Knecht-Hermann

DATE

Date JUL 7 '61 S. O. 4678

MOUNTED

747622

JAN 11 1961

Temp. Coef. = — Microvolts per °C

B. S. Certified Value

1.01 Volts at °C

Eppley Certified Value JUL 7 '85

1.01 924 Abs. Volts at 24 °C

Shipped to Braun-Koeck-Hiemann

Calif

Date JUL 7 '81 S. O. 4678