

**Equipment Information:**

Job No.:  
Manufacturer: KEITHLEY  
Description: MULTIMETER, DIGITAL  
Department:  
Temp./RH: 21.0 C / 41.0 %  
Cal. Interval: 12 MONTHS  
As Found: IN TOLERANCE  
Cal Date: 07/27/2017  
Owner:

**Purchase Order: CREDIT CARD**

Asset Tag No.: SE-0001  
Model Number: 2002  
Serial Number: 0727052  
Inspected By: XXXXXXXXXX  
Job Title: METROLOGIST  
Calibration Result: PASSED  
As Left: IN TOLERANCE  
Cal. Due Date: 07/27/2018  
Owner's Email:

**Procedure: KEITHLEY\_2002****Calibration Notes:**

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	As Left	UOM
1	DCV, 200mV RANGE	100.00000	99.99630	100.00370	99.99713	99.99713	DCmV
2	DCV, 2V RANGE	1.0000000	0.9999882	1.0000118	0.9999951	0.9999951	DCV
3	DCV, 20V RANGE	10.000000	9.999897	10.000103	9.999992	9.999992	DCV
4	DCV, 20V RANGE	-10.000000	-10.000103	-9.999897	-9.999994	-9.999994	DCV
5	DCV, 200V RANGE	100.00000	99.99740	100.00260	100.00048	100.00048	DCV
6	DCV, 1000V RANGE	1000.0000	999.9760	1000.0240	1000.0145	1000.0145	DCV
7	ACV, 1KHz, 200mV RANGE	100.000	99.960	100.040	99.960	99.960	ACmV