

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	As Left	UOM
21	DCA, 2mA RANGE	1.000000	0.999610	1.000390	0.999917	0.999917	DCmA
22	DCA, 20mA RANGE	10.00000	9.99610	10.00390	9.99964	9.99964	DCmA
23	DCA, 200mA RANGE	100.0000	99.9585	100.0415	100.0055	100.0055	DCmA
24	DCA, 2A RANGE	1.000000	0.999210	1.000790	0.999947	0.999947	DCA
25	ACA, 1KHz, 200uA RANGE	100.000	99.570	100.430	99.902	99.902	ACuA
26	ACA, 1KHz, 2mA RANGE	1.00000	0.99850	1.00150	0.99948	0.99948	ACmA
27	ACA, 1KHz, 20mA RANGE	10.0000	9.9850	10.0150	9.9965	9.9965	ACmA
28	ACA, 1KHz, 200mA RANGE	100.000	99.820	100.180	100.006	100.006	ACmA
29	ACA, 1KHz, 2A RANGE	1.00000	0.99670	1.00330	0.99876	0.99876	ACA
30	FREQUENCY	10.0000	9.9970	10.0030	10.0000	10.0000	KHz

#### Standards Used To Calibrate Equipment

I.D.	Manufacturer	Model	Description	Serial Number	Cal. Due Date
00009988	FLUKE	5720A	CALIBRATOR	6740202	08/16/2017