

Certificate of Calibration



Standards Laboratory Calibration
Certificate Number WO-0045



Model Number 732B
Manufacturer Fluke Corp
Description Standard
Serial Number

Date of Calibration 12 Jun 2023
Procedure 24BR-732ABC-R2
Temperature (23 ± 1) °C
Humidity (45 ± 10) %RH

Customer

.....
United States

Location of Calibration

Keysight Technologies Inc.
10090 Foothills Blvd.
Roseville CA 95747-7102
UNITED STATES

This certifies that the equipment has been calibrated using applicable Keysight Technologies procedures and accredited to ISO/IEC 17025:2017 and ANSI/NCSL Z540.3-2006. The quality management system is registered to ISO 9001:2015.

As Received Conditions

The measured values of the equipment was for the purpose of characterization. No compliance statement is made relating to specification.

Action Taken

- No corrective actions were necessary.

As Completed Conditions

The measured values of the equipment was for the purpose of characterization. No compliance statement is made relating to specification.

The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%. This probability corresponds to a coverage factor of k=2 for a normal distribution.

Remarks or Special Requirements

This calibration report shall not be reproduced, except in full. The documented results relate to the equipment calibrated only.

The test limits stated in the report correspond to the published specifications of the equipment, at the points tested. This calibration report may refer to equipment manufactured by HP, Agilent and Keysight as being manufactured by Keysight Technologies. This calibration report is available at Keysight Support Portal <http://support.keysight.com/>.

Keysight Technologies Inc
10090 Foothills Blvd.
Roseville CA 95747-7102
UNITED STATES

Wes Fischbach Roseville Serv. Cntr. Mgr.