

## No upgrade between bold lines!!!!

This will damage your instrument DG1022 seriously.

Rigol Technologies EU will charge the repair of defects caused by wrong update.

数字板	数字FPGA	DSP
00.02	00.06	00.02.07.00.05
00.03	00.08	00.02.08.00.01
		00.02.10.00.03
	00.09	00.02.11.00.00

Example 1:

Current firmware version is **00.03.00.08.00.02.08.00.01**

This unit **can not** be upgraded to **00.03.08.00.02.10.00.03** and

**Can also not** be upgraded to latest version **00.03.09.00.02.11.00.00**

Example 2:

Current firmware version is **00.03.08.00.02.10.00.03**

This unit **can** be upgraded to the latest version **00.03.09.00.02.11.00.00**