



Updating the SDM3000 firmware

Insert USB flash drive into the SDS3000 meter and press [Shift] and [2nd] to enter the Setting interface. Select [Test/Admin], enter the secondary interface to select [Firmware Update]. Press [Browse] to browse the file directory and select the new flash drive firmware (.ADS file). Select [Update] to begin the installation. After completion, restart the SDM3000.